

Certificate of Analysis

Description: Corning® glutagro™ Supplement

200 mM L-alanyl-L-glutamine

Storage: 2-30°C

Catalog #: 25-015-C1

Lot #: 17216008

Expiration Date: 06/18

Country of Manufacture: USA

Notes:

Mediatech, Inc. products are prepared by a validated aseptic sterile filtration process to ensure that all products comply with the industry's Sterility Assurance Level (SAL) of at least 10^{-3} (i.e. demonstrating that a product produced in this manner has no more than 1 random contaminant per 1000 units).

This product was tested in accordance with currently approved Mediatech, Inc. specifications and procedures. Testing procedures are maintained in compliance with the current versions of the USP and/or EP, where applicable.

For Research Use Only. Not for Human Therapeutic Use. Utilization of this product apart from the labeled intended use may be a violation of local and/or Federal Law.


TEST PARAMETERS	SPECIFICATION	RESULT
pH	5.5-5.9 @ 20-25°C	5.7 @ 20°C
Osmolality	438-498 mOsm/Kg H ₂ O	468 mOsm/Kg H ₂ O
Endotoxin	< 0.25 EU/mL	0.046 EU/mL
Mycoplasma	Tested Negative	Negative
Sterility	Pass	Pass
Cell Line		
AE-1 (Cellgrowth)	Pass	Pass
MRC-5 (Cellgrowth)	Pass	Pass

For additional information on this Mediatech product, please contact our Technical Services department at (800) 235-5476 or via email at MTTechServ@corning.com.

Following signatures indicate the above material has met all quality specifications and has been reviewed by a Quality representative.

Written By/Date: Eugene Jaimes 11 Jul 2016

Reviewed By/Date: George 11 JUL 2016



Mediatech, Inc.
A Corning Subsidiary
9345 Discovery Boulevard
Manassas, VA 20109
Main (703) 471-5955
Customer Service (800) 235-5476
www.cellgro.com