

Corning Incorporated - Life Sciences
271 County Route 64
Big Flats NY 14814 USA
www.corning.com/lifesciences
Refer to website for regional contact information.

Page: 1 / 1

Product Name	: CultureSlide 4well 1.7cm2 24cs		
Catalog Number	: 354114	Manufacture Date	: 2018-10-02
Lot ID	: 1002184000		
Expiration Date	: 2020-10-31		

Quality Management System - Complies with the current version of the ISO 9001 Standard.

USP Class VI Testing - All material resin is tested, qualified and shown to be non-toxic as established in the Standards USP Class VI Chapter<87>, "Biological reactivity Tests, in Vitro" and Chapter<88>, "Biological Reactivity Tests, in vivo".

Tissue Culture - Tested for the attribute of cell attachment and growth utilizing an attachment-dependent mammalian cell line. A minimum of 95% confluency is required for acceptance.

Sterility - Product has been sterilized and parametrically released per the requirements of ANSI/AAMI/ISO 11135, "Sterilization of health care products - Ethylene oxide". Products meet a minimum Sterility Assurance Level (SAL) of 10^{-6} .

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release.

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



Edwin A Kindler
Div. Quality Engineering Manager