

**Corning Incorporated
Life Sciences**

Registered
ISO 9001:2008

Certificate of Compliance

Catalog Number: 3737

Lot Number: 13514007

Product Description: Costar® 12-well non treated cell culture multiple well plate, flat bottom with lid

Component Materials:
 Plate/Lid - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.

Sterilization:
 This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.
 Sterility Assurance Level: SAL 10⁻³

Pyrogens:
 The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 *Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing*.
 Results: ≤ 0.1 EU/mL (≤ 4EU/device)


Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:
 This product is manufactured with animal free materials.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

| | |
|------------------------|-----------------|
| Test/Procedure: | Results: |
| No Surface Treatment | Pass |
| Visual Attributes | Pass |
| Packaging | Pass |

Lot Number Designation:
 8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Expiration Statement:
 This product has an expiration date of 3 years from the date of manufacture for the attributes contained within this certificate.



Judith A. Gibbs
 Quality Assurance Manager

Rev No: 5