

CORNING LIFE SCIENCE (WUJIANG) CO LTD - Life Sciences
T03/17 BUILDING, 1801 PANG JIN ROAD
WUJIANG, JIANGSU PROVINCE 215200 CHN
www.corning.com/lifesciences
Refer to website for regional contact information.

Page: 1 / 1

Product Name	: TUBE,CENT,15ML,PLG, PP,BULK	Manufacture Date	: 2021-05-26
Catalog Number	: 430766		
Lot ID	: 14621601		
Expiration Date	: 2024-05-25		

Quality Management System - Complies with the current version of the EN ISO 13485 Standard.

BSE/TSE - Product complies with the latest revision of EMA/410/01 "Note for Guidance on minimising the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

Non-Pyrogenic - Tested and met the criteria established in the current version of ANSI/AAMI ST 72, "Bacterial Endotoxins - Test methodologies, routine monitoring, and alternatives to batch testing" and USP <85>, "Bacterial Endotoxins Test". The acceptance level for product is ≤ 0.10 EU/ml or ≤ 4 EU/device.

USP Class VI Testing - All material resin is tested, qualified and shown to be non-toxic as established in the Standards USP Class VI Chapter<87>, "Biological reactivity Tests, in Vitro" and Chapter<88>, "Biological Reactivity Tests, in vivo".

Sterility - Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, "Sterilization of health care products- Radiation". Products meet a minimum Sterility Assurance Level (SAL) of 10^{-5} .

Sterility - Products labeled Sterile Fluid Path have been designed to ensure sterility of the portion of the product intended for contact with fluids.

DNase/RNase Free - Tested by nuclease assay method and found to be free of detectable DNase/RNase contamination.

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

- Visual Inspection - Pass
- Functional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.

Thomas Liu

Huihao (Thomas) Liu
Wujiang Quality Manager