

CORNING GOSSELIN SAS  
123 Rue de Caëstre - CS40019  
59529 HAZEBROUCK CEDEX  
BORRE - FRANCE

Date of certificate : 2018 - 11 - 14

Page : 1/1

**Product description**

*Octagonal Bottle, 250 ml PET nagrauated H142 Ø31 \$ tamper-evident screw cap na, 24/filmed pack, S*

**Catalogue number**

431732

**Lot number**

677578

**Sanitary Assurance Level**

*This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10<sup>-6</sup>*

**Manufacturing date**

2018/04/17

**Expiration date**

2021/04

**Expiration statement**

*This product is designed to have and expiration date of 3 years from the date of manufacture regarding sterility.*

**Performance testing (AQL)**

*Inspection based on a statistically significant number of samples towards to internal or agreed specifications that determine Assurance Quality Level.*



Cédric MAES  
Plant Quality Manager, Borre & Chorges  
maesc@corning.com