

## Certificate of Analysis

**Description: Molecular Biology Grade Water**

Tested to USP Sterile Purified Water Specifications

**Catalog #:** 46-000-CI

**Lot #:** 23217005

**Expiration Date:** 08/19

**Storage:** 2-30°C

**Country of Manufacture:** USA

**Notes:**

Mediatech, Inc. products are prepared by a validated aseptic sterile filtration process to ensure that all products comply with the industry's Sterility Assurance Level (SAL) of at least  $10^{-3}$  (i.e. demonstrating that a product produced in this manner has no more than 1 random contaminant per 1000 units).

This product was tested in accordance with currently approved Mediatech, Inc. specifications and procedures. Testing procedures are maintained in compliance with the current versions of the USP and/or EP, where applicable.

For Research Use Only. Not for Human Therapeutic Use. Utilization of this product apart from the labeled intended use may be a violation of local and/or Federal Law.

| TEST PARAMETERS                             | SPECIFICATION                 | RESULT                   |
|---|-------------------------------|--------------------------|
| Sterility USP                               | Pass                          | Pass                     |
| pH  | 6.0 ± 1.0 @ 20-25°C           | 5.5 @ 25°C               |
| Appearance                                  | Clear colorless solution      | Clear colorless solution |
| Conductivity (before packaging), USP        | Test and Report               | 0.24 µS/cm @ 85°C        |
| Conductivity (after packaging)              | Meets USP Requirements        | 2 µS/cm @ 26°C           |
| Total Organic Carbon (before packaging) USP | Test and Report               | 12 ppb                   |
| Oxidizable Substances                       | Solution remains faintly pink | Pass                     |
| RNase                                       | None Detected                 | None Detected            |
| DNase                                       | None Detected                 | None Detected            |
| Protease                                    | None Detected                 | None Detected            |

For additional information on this Mediatech product, please contact our Technical Services department at (800) 235-5476 or via email at [MTTechServ@corning.com](mailto:MTTechServ@corning.com).

Following signatures indicate the above material has met all quality specifications and has been reviewed by a Quality representative.

Written By/Date: *[Signature]* 22 Aug 2017

Reviewed By/Date: *[Signature]* / 4E  
22 Aug 2017



Mediatech, Inc.  
A Corning Subsidiary  
9345 Discovery Boulevard  
Manassas, VA 20109  
Main (703) 471-5955  
Customer Service (800) 235-5476  
[www.cellgro.com](http://www.cellgro.com)