

**Corning Incorporated  
Life Sciences**

Registered  
ISO 9001:2008

# Certificate of Compliance

**Catalog Number:** 4806  
**Lot Number:** 15314003  
**Product Description:** Corning® Universal Fit Pipette Tip, 1 - 200 µL, 480 tip rack

- Component Materials:**
- Tip - Virgin Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.
  - Rack - Polypropylene with orange colorant, meets *USP, Class VI* requirements for plastic containers and closures. Heavy metal free (meets *CONEG* req.) color concentrate.
  - Base - Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.

**Special Feature:**  
The tips are autoclavable.

**RNase/DNase Testing:**  
This product has been tested and is free of any detectable RNase/DNase contamination.

**Pyrogens:**  
The product has been tested and has met the criteria established in ANSI/AAMI ST 72:2002/(R) 2010: *Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing*. Results: less than 0.1 EU/mL

**Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:**  
This product complies with the latest revision of EMEA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicine products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/410/01.

This manufacturing lot has been sampled and tested in accordance with Standard Operating Procedures and has been released by Quality Assurance for the following characteristics:

<b>Test/Procedure:</b>	<b>Results:</b>
Visual Attributes	Pass
Packaging	Pass

**Lot Number Designation:**  
8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.



Viviana Niño  
Quality Supervisor

Rev No: 5