

CORNING GOSSELIN SAS
123 Rue de Caëstre - CS40019
59529 HAZEBROUCK CEDEX
BORRE - FRANCE

Date of certificate : 2018 - 11 - 16

Page : 1/1

Product description

Bottle Oct PET 125 mL 31mmTECap Non-assembled 24/Pack 288/CS

Catalogue number

P125B-05

Lot number

680573

Sanitary Assurance Level

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10⁻⁶

Manufacturing date

2018/05/17

Expiration statement

This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards internal or agreed specifications that determine Assurance Quality Level.



Cédric MAES
Plant Quality Manager, Borre & Chorges
maesc@corning.com