

CORNING GOSSELIN SAS
123 Rue de Caëstre - CS40019
59529 HAZEBROUCK CEDEX
BORRE - FRANCE

Date of certificate : 2018 - 10 - 02

Page : 1/1

Product description

Container 60mL PP GreenScrewCap Sterile 70/Bag 700/CS

Catalogue number

TP35C-004

Lot number

619482

Sanitary Assurance Level

Irradiated. Finished products with an irradiated claim are ionised and not subjected to sterilization process according ISO 11137-2. They therefore do not make a claim to have a defined Sterility Assurance Level.

Manufacturing date

2016/02/03

Expiration date

2019/02

Expiration statement

This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards internal or agreed specifications that determine Assurance Quality Level.



Cédric MAES
Plant Quality Manager, Borre & Chorges
maesc@corning.com