

CORNING GOSSELIN SAS  
123 Rue de Caëstre - CS40019  
59529 HAZEBROUCK CEDEX  
BORRE - FRANCE

Date of certificate : 2018 - 11 - 14

Page : 1/1

**Product description**

*Container 60mL PP GreenScrewCap Sterile 70/Bag 700/CS*

**Catalogue number**

*TP35C-004*

**Lot number**

*659470*

**Sanitary Assurance Level**

*Irradiated. Finished products with an irradiated claim are ionised and not subjected to sterilization process according ISO 11137-2. They therefore do not make a claim to have a defined Sterility Assurance Level.*

**Manufacturing date**

*2017/08/07*

**Expiration date**

*2020/08*

**Expiration statement**

*This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.*

**Performance testing (AQL)**

*Inspection based on a statistically significant number of samples towards internal or agreed specifications that determine Assurance Quality Level.*



Cédric MAES  
Plant Quality Manager, Borre & Chorges  
maesc@corning.com