

CORNING GOSSELIN SAS
123 Rue de Caëstre - CS40019
59529 HAZEBROUCK CEDEX
BORRE - FRANCE

Date of certificate : 2020 - 06 - 10

Page : 1/1

Product description

Container 125mL PP RedScrewCap Ster ile 380/CS

Catalogue number

TP52C-002

Lot number

743073

Manufacturing date

2020/03/11

Sterility Assurance Level

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10⁻³

Expiration date

2023/03/01

Expiration statement

This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.

Performance testing (AQL)

Inspection based on a statistically significant number of samples towards internal or agreed specifications that determine Assurance Quality Level.



Cédric MAES
Plant Quality Manager, Borre
maesc@corning.com