

CORNING GOSSELIN SAS  
123 Rue de Caëstre - CS40019  
59529 HAZEBROUCK CEDEX  
BORRE - FRANCE

Date of certificate : 2021 - 07 - 19

Page : 1/1

### **Product description**

*Container 125mL PP RedScrewCap Ster ile 380/CS*

### **Catalogue number**

*TP52C-002*

### **Lot number**

*790396*

### **Manufacturing date**

*2021/06/22*

### **Sterility Assurance Level**

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of  $10^{-3}$

### **Expiration date**

*2024/06/01*

### **Expiration statement**

*This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.*

### **Performance testing (AQL)**

*Inspection based on a statistically significant number of samples towards internal or agreed specifications that determine Assurance Quality Level.*



Hugues Vanderschelden  
Plant Quality Manager, Borre  
vandersch2@corning.com