

CORNING GOSSELIN SAS  
123 Rue de Caëstre - CS40019  
59529 HAZEBROUCK CEDEX  
BORRE - FRANCE

Date of certificate : 2021 - 06 - 11

Page : 1/1

### **Product description**

*Container 180mL PP BlueScrewCap Ste rile 264/CS*

### **Catalogue number**

*TP53C-001*

### **Manufacturing date**

*2021/05/05*

### **Lot number**

*785156*

### **Sterility Assurance Level**

Irradiated. Finished products with an irradiated claim are ionised and not subjected to sterilization process according ISO 11137-2. They therefore do not make a claim to have a defined Sterility Assurance Level.

### **Expiration date**

*2024/05/01*

### **Expiration statement**

*This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.*

### **Performance testing (AQL)**

*Inspection based on a statistically significant number of samples towards internal or agreed specifications that determine Assurance Quality Level.*



Cédric MAES  
Plant Quality Manager, Borre  
maesc@corning.com