

CORNING GOSSELIN SAS  
123 Rue de Caëstre - CS40019  
59529 HAZEBROUCK CEDEX  
BORRE - FRANCE

Date of certificate : 2020 - 09 - 07

Page : 1/1

### Product description

*Container 180mL PP RedScrewCap Ster ile 264/CS*

### Catalogue number

*TP53C-002*

### Lot number

*746066*

### Manufacturing date

*2020/04/07*

### Sterility Assurance Level

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10<sup>-3</sup>

### Expiration date

*2023/04/01*

### Expiration statement

*This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.*

### Performance testing (AQL)

*Inspection based on a statistically significant number of samples towards internal or agreed specifications that determine Assurance Quality Level.*



Cédric MAES  
Plant Quality Manager, Borre  
maesc@corning.com