

CORNING GOSSELIN SAS  
123 Rue de Caëstre - CS40019  
59529 HAZEBROUCK CEDEX  
BORRE - FRANCE

Date of certificate : 2018 - 10 - 09

Page : 1/1

**Product description**

*Container 180mL PP GreenScrewCap Sterile 264/CS*

**Catalogue number**

*TP53C-004*

**Lot number**

*646220*

**Sanitary Assurance Level**

*This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10<sup>-3</sup>*

**Manufacturing date**

*2017/02/10*

**Expiration date**

*2020/02*

**Expiration statement**

*This product is designed to have an expiration date of 3 years from the date of manufacture regarding sterility.*

**Performance testing (AQL)**

*Inspection based on a statistically significant number of samples towards internal or agreed specifications that determine Assurance Quality Level.*



Cédric MAES  
Plant Quality Manager, Borre & Chorges  
maesc@corning.com